

# Certificate

## Quantitative test method for the evaluation of bactericidal activity on non-porous surfaces without mechanical action against *Enterococcus hirae*

Herewith we certify, that the laboratory of

**bactologicum GmbH**  
**Fraunhoferstr. 3**  
**D 25524 Itzehoe**

participated successfully  
in the VAH ring trial R2022-01

Remarks on the trial design:

- The proficiency test took place from 30 May 2022 to 29 July 2022
- Each laboratory performed the test according to VAH method 14.1; DIN EN 17387
- The test design included the following:
  - < Test organism: *Enterococcus hirae*
  - < Test product: Product A
  - < Interfering substances: dirty conditions
  - < Concentration-time-ratio of product A: 30% - 5 min; 40% - 5 min; 50% - 5 min; 60% - 5 min
  - < Active ingredients of product A: Ethanol

This certificate is only valid in combination with the evaluation report of VAH ring trial R2022-01

Bonn, 13 September 2022



Dr. rer. nat. S. Gemein



Dr. rer. nat. J. Gebel

Verbund für Angewandte Hygiene e.V. (VAH)

