

# Certificate

## Quantitative suspension test for the evaluation of bactericidal activity in the medical area under dirty conditions against *Pseudomonas aeruginosa*

Herewith we certify, that the laboratory of

**bactologicum GmbH**  
**Fraunhoferstr. 3**  
**D- 25524 Itzehoe**

participated successfully

in the VAH ring trial R2026-01

Remarks on the trial design:

- The proficiency test took place from 02 February 2026 to 23 March 2026
- Each laboratory performed the test according to DIN EN 13727 or VAH 9
- The test design included the following:
  - < Test organism: *Pseudomonas aeruginosa*
  - < Test product: product A
  - < Interfering substances: dirty condition
  - < Concentration-time-ratio of product A: 1 % - 5 min/ 15 min/ 60 min
  - < Active ingredients of product A: amine-based

This certificate is only valid in combination with the evaluation report of VAH ring trial R2026-01

Bonn, 04 May 2026



K. Roesch M.Sc.



Dr. rer. nat. J. Gebel

Verbund für Angewandte Hygiene e.V. (VAH)

